

No. 2013-017369-01



# 안전인증서

## Certificate of Safety

**GASTRON Co.,LTD.**

75-10 Palgok 2-dong, SangRok-gu, Ansan-city Kyunggi-Do, Korea

*The following product manufactured by the above company has been authorized for the use of safety certification mark as the product has satisfied the safety and health requirements under Article 34 of Occupational Safety and Health ACT and paragreaph 4 of Article 58-4 of Enforcement Ordinance of the same ACT.*

### Electrical Equipment

GAS DETECTOR

### Model (Marking) / Certificate No.

GTD-5000F (Ex d IIC T6 IP65) / 13-KB2BO-0448

### Applied standards

Announcement No. 2010-36 of Ministry of Employment and Labor

### Conditions of Certification

- 1. Manufacturing location(s)** : This certificate is only valid for the specified products produced at 75-10 Palgok 2-dong, SangRok-gu, Ansan-city Kyunggi-Do, Korea.
- 2. Description of equipment** : GAS DETECTOR  
·Rating : 24VDC, Max. 600mA
- 3. Scope of certification** : This certificate is only valid for the models as above.
- 4. Special conditions for safe use**  
·Ambient Temperature : -40℃ ~ +60℃
- 5. Details of certificate changes** : None
- 6. Details of the others** : See Instruction manual

Date of Issue : 2013-8-21

**한국산업기술시험원장**

**KOREA TESTING LABORATORY**

